

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Harvey Cedars P.D., Harvey Cedars, NJ

Date Received: 01/30/2018

Date given to service: 01/30/2018

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR UL-0049

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	N/C
4414166	Labor	.5	N/C
6808455	Fuel Cell	1	N/C

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARKL-0987

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/21/2018