

# RETURN & REPAIR FORM



*NJ*

CUSTOMER MANTUA Twp PD Cust. # \_\_\_\_\_  
 Date ret. 5/10 Carrier Fed Exp Method 3PM RMA# \_\_\_\_\_  
 7110 Serial# ARJL-0070 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
<i>w</i>   <u>6810269</u>	<u>ESCAP MOTOR</u>		<u>1</u>	
<i>w</i>   <u>MPCAL 71</u>			<u>1</u>	
		Repair Time		Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
Replaced ESCAP MOTOR, Calibrated UNIT full PC  
⊕ OPS CHK. UNIT HAS VERSION NJ-3.11.

Date 11 MAY 05 Technician [Signature]  
 Every above checked item is included  Shipping \_\_\_\_\_