

RETURN AND REPAIR FORM



Customer Information

B: \_\_\_\_\_ S: \_\_\_\_\_

Company Name: Mantua Twp P.D. Mantua, NJ

Date Received: 10/20/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810 Serial Number: AR UH - 0070  
 7110  8610 Printer Serial#: AR  
 7510  DT5000 Sim Serial#: DD  
 7410 Upper-half  9510 Probe Serial#: DD  
 7410 Whole

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>mp c/c 71</u>	<u>Calibration</u>	<u>1</u>	<u>n/c w</u>
<u>mp labor</u>	<u>labor</u>	<u>1.5</u>	<u>n/c w</u>

Repair Notes: Replaced capacitor at located at position C-162.

cal. w/DC; ops check

Service Technician: [Signature] Date: 10-21-11