

# Return & Repair Form:

## Customer Information:

Company Name: NEW JERSEY STATE PD B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 1/6/10 Date Given to Service: 1/6/10

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: ARUL-0075  
 9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top  $\downarrow$  USED PCB \_\_\_\_\_  
 Other \_\_\_\_\_ FROM DEMO \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter  Reg 7110 ST# \_\_\_\_\_  
 Printer Paper  Pri ARM-0607 \_\_\_\_\_  
 Mouthpieces  Car \_\_\_\_\_

Other (Please Specify) \_\_\_\_\_

## Repair Information:

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>6315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C W/ARM</u>
<u>MP CAL 41</u>	<u>CAL</u>	<u>1</u>	<u>N/C W/ARM</u>
<u>MP Labor</u>	<u>Labor</u>	<u>0.5</u>	<u>N/C W/ARM</u>

### Repair Notes:

INSTALL PCB FROM DEMO UNIT ARM-0607

Cal w/Qc & Ops Check

Service Technician: AS

Date: 01/08/2010