

Return & Repair Form

Customer Information:

CUSTOMER GIBBSBORO BORO P.D. Cust #
Date of Return: 1/23/2008 Carrier: (FX) UPS DHL USPS Method: GRD (NDA) 2DY 3DY OTHER: STAD

Product Information:

Product: (7110) 7410 6510 6810 Serial # ARUM - 0061
Description: A B PLUS SCREENER Printer Serial # AR -
Whole Inst. Top Half Simulator Serial #
OTHER:
Probe Serial # DD P -

Demo Unit Returned to stock on:

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr #
Printer Paper Printer Ribbon Casio #
Mouthpieces Carrying Case Dry Gas
Other (specify):

Warranty Expiration Date:

Repair Information:

Test #

Reason for Return: HIGH IR RESULTS

Table with 4 columns: Part #, Description, Qty, Total. Rows include 6808455 FC, MPCAL71 7110 CMC, and LABOR LABOR.

Repair Notes: REPLACED FC-WAS UJ SERIES / REPLACED IR DETECTOR PCB (PN 680945B) NON-STOCK PART - DUE TO IR RESULTS HIGH ON CUSTOMER TESTS PROVIDED / CHECK VALVE OK /

Technician: [Signature]

Date: 1-24-08