

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: NJSP Bass River, New Greton NJ

Date Received: 03/27/12 Date given to service: 03/27/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR UM-0062

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>labor</u>	<u>1</u>	<u>N/C W</u>

Repair Notes: Replaced capacitor at location R-156 to restore modem ethernet connection.

cal w/DC & ops check

Service Technician: [Signature]

Date: 3-30-12