

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER BOND Brook PD Cust # 56340
 Date of Return: 2/6 Carrier: Fed Exp Method: STD

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWA-0169
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Error (032) Motor, Switch

Part #	Description	Qty	Total
<u>6810269</u>	<u>EC Motor</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced EC Motor because it wouldn't turn. Cal ops check + QC.

Technician: CRD

Date: 2/8/06