

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Bound Brook P.A., Bound Brook, N.J.

Date Received: 02/27/12 Date given to service: 02/27/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WA-0169  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP CCL 71	Calibration	1	N/C w
6808455	Fuel cell	1	N/C w
6808426	Plate	1	N/C w
mp Labor	Labor	.5	N/C w

Repair Notes: Replaced busted Fuel Cell # ARZK-0734 with Fuel Cell # ARCK-0889. The unit was sent in because it would not warm up, I replaced the cuvette Heater Fuse.

cal w/OC; ops check

Service Technician: J.S Date: 3-19-12