

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Bound Brook P.D., Bound Brook, NJ

Date Received: 03/29/2016 Date given to service: 03/29/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WA-0169  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: January 2016 (56340)

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell.  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Fuel Cell = ARJB-1455  
 CAL W/QC AND OPS CHECK

Service Technician: BC Date: 04/11/2016