

# Return & Repair Form:

## Customer Information:

Company Name: NJSP B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 12-27-2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other \_\_\_\_\_

## Product Information:

Product: 7410 7110 - 6510 - 6810 - 7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: ARWA-0171  
 Printer Ser #: AR  
 Whole Instrument \_\_\_\_\_ Top 1/2 \_\_\_\_\_  
 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Reason for Return:		Test #	
Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>n/c w</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>MP labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

### Repair Notes:

PCT ARZK-0638 Installed 12-01-08  
NEW Captive  
Cal w/DC; ops check

Service Technician: [Signature]

Date: 1/28/2011