

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER: BOROUGH OF Point Pleasant Cust # _____
 Date of Return: 2/8 Carrier: FedEx Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # PRWA-0175
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Wont power up

Part #	Description	Qty	Total
12029	Power Supply	1	N/C
MPLABOR	LABOR	.5	N/C
MPCAL71	7110 Cal	1	

Repair Notes: Replaced dead power supply. Cal ops check & O.C.

Technician: CPD

Date: 2/9/06