

## Return & Repair Form

### Customer Information:

CUSTOMER PT. Pleasant NJ Cust # 56604  
 Date of Return: 10/9/06 Carrier: FedEx Method: Exp. Saver

### Product Information:

Product : 7110 7410 6510 6810 Serial # ARWA-0175  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DDUNP2-293  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: ERROR 071 & 023 / Probe defective

Part #	Description	Qty	Total
<u>12025</u>	<u>Probe</u>	<u>1</u>	
<u>MPQAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Probe failed O.C. Replace + return to stock  
Replacement probe serial # is DDXAP2-020.

Could not duplicate error 071 or 023. Full cal ops  
check + O.C.

Technician: CRD

Date: 10/11/06