

Return & Repair Form:

Customer Information:

Company Name: Point Pleasant Beach B: _____ S: _____
 Date Received: 1-26-11 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6310-6810-7510
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo
 Serial #: AR 107-0175
 Printer Ser #: AR -
 Whole Instrument _____ Top 1/2 _____
 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:		Test #		
<u>FC</u>				
Part Number	Description	Pr	Qty	Total Cost
<u>6808455</u>	<u>FC</u>		<u>1</u>	<u>n/cw</u>
<u>6808486</u>	<u>Plate</u>		<u>1</u>	<u>n/cw</u>
<u>mp cal 71</u>	<u>cal</u>		<u>1</u>	<u>n/cw</u>
<u>mp labor</u>	<u>labor</u>		<u>.5</u>	<u>n/cw</u>

Repair Notes:
FC# ARUN-0043 replaced with FC#
AR 13L-1860.
cal w/DC ; ops check

Service Technician: [Signature]

Date: 1/31/2011