

# Return & Repair Form:

## Customer Information:

Company Name: Forward P.O. B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 6-9-09 Date Given to Service: \_\_\_\_\_

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other

## Product Information:

Product: 7410 7110 6510 - 6810 Serial #: AR WA - 0176

Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808435</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP LABOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:  
REPLACE FC - CAL W/ - QC & OPS CHECKS.  
REPLACED 1 FC PLATE

Service Technician: SC

Date: 06-10-09