

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Bordertown PD Cust # _____
 Date of Return: 11/3 Carrier: FedEx Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # FRWA-0179
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: IR Problem

Part #	Description	Qty	Total
	MISC PART	1	N/C
MPCAL71	M110 CAL	1	N/C

Repair Notes: D81 IC on Mother board was defective.
Replaced D81. Cal ops check + Q.C.

Technician: Aut Dely

Date: 11/7/05

Handwritten initials/signature