

# Return & Repair Form:

## Customer Information:

Company Name: Borden Town Twp B: 57208

Date Received: 11/19/2010 Date Given to Service: \_\_\_\_\_

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WA - 0179

9510-Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR

Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_ Probe Ser # DD P

## ACCESSORIES:

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_

Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_

Mouthpieces  Carrying Case  Dry Gas

Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date Oct. 2010

## Repair Information:


Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>n/c w</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>n/c w</u>
<u>MP Cal 41</u>	<u>Cal</u>	<u>1</u>	<u>n/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

## Repair Notes:

FLT ARBH-1222

Cal w/ QC i ops check

Service Technician: 

Date: 11/22/2010