

Return & Repair Form

Customer Information:

CUSTOMER: CITY OF SUMMIT, NJ Cust # \_\_\_\_\_  
 Date of Return: 8/13/07 Carrier: \_\_\_\_\_ Method: \_\_\_\_\_

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWA - 0180  
 Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # \_\_\_\_\_
- Printer Paper
- Printer Ribbon
- Casio # \_\_\_\_\_
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): POWER CORD

Warranty Expiration Date: \_\_\_\_\_

Repair Information:

Test # \_\_\_\_\_  
 Reason for Return: EC RESULTS = 0.000 ON 10% STD CHECKS

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARRANTY</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	
<u>MPLCAL71</u>	<u>7110 CAL</u>	<u>1</u>	
<u>MPLLABOR</u>	<u>LABOR</u>	<u>.5</u>	

Repair Notes: CORROSION OF FC / REPLACED W/ "YD" / REPLACED FC PLATE / CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 08-14-07