

# Return & Repair Form:

## Customer Information:

Company Name: Paulsboro Twp B: 57153 S: \_\_\_\_\_  
 Date Received: 6-21-10 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other \_\_\_\_\_

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WFA - 0183  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date JULY 2010

## Repair Information:

Test #

Reason for Return: FC / motor

Part Number	Description	Qty	Total Cost
<u>6808374</u>	<u>motor</u>	<u>1</u>	<u>N/C W</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>cal</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>N/C W</u>

## Repair Notes:

Installed new FC AR130-2531, and new motor.

cal w/ QC & ops check

Service Technician: 

Date: 06/21/2010