

# Dräger safety

## Return & Repair Form

### Customer Information:

CITY ✓  
 CUSTOMER Vinland P D Cust # 57094  
 Date of Return: 11/28/2007 Carrier: EX UPS DHL USPS Method: GRD NDA 2DY 3DY  
 OTHER: \_\_\_\_\_

### Product Information:

Product : 7110 7410 6510 6810 Serial # AR WA - 0185  
 Description: A B PLUS SCREENER \_ Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: Cal Cert

Part #	Description	Qty	Total
<del>45470</del>			
<u>6808455</u>	<u>FUELCELL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR N/C</u>

Repair Notes: REPLACE FC / CAL w/ QC & OPS CHECKS

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

Technician: [Signature]

Date: 11-29-07