

# Return & Repair Form

## Customer Information:

Company Name: Vineland Police Dept. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 8-1-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR WA - 0185  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: ERROR 32

Part Number	Description	Qty	Total Cost
<u>6810269</u>	<u>MOTOR</u>	<u>1</u>	<u>N/CWARR</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/CWARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/CWARR</u>

## Repair Notes:

ERROR 32 DUPLICATED / REPLACE MOTOR /  
7110 WAS HERE IN NOV '07 FOR FC REPLACEMENT  
CAL W/ QCE OPS CHECKS.

Service Technician: 08-04-08

Date: [Signature]