

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Cransford Twp, P.D., Cransford, NJ

Date Received: 10/01/13

Date given to service: 10/01/13

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WA 0187

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: July 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	4	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell, spacer plates, cermetek chip, c156, and c167.

Complaint that the unit would not connect to the modem and this issue was due to bad C156 and C167.

Fuel Cell = AREF-1698

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/10/13