

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Montville Twp. Police Dept. Cust # 37347

Date of Return: 11/20/06 Carrier: FedEx Method: 2-Day

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWA-0191

Description: A B PLUS SCREENER _ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER : _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Low SOLUTION - CHANGES

Part #	Description	Qty	Total
<u>6809S12</u>	<u>Cuvette check valve</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced check valve in cuvette because it was stuck. Full cal ops check + QC.

Technician: CRD

Date: 11/21/06