

Return & Repair Form: 10457071, 80578604

Customer Information:

Company Name: Mansfield Twp B: 57359 S:
Date Received: 2-7-11 Date Given to Service:
Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WA- 0193
9510- Drug Tester 5000
Description: A - B - Plus - Screener - Demo Printer Ser #: AR
Whole Instrument Top 1/2 Sim Ser #:
Other Probe Ser # DD P

ACCESSORIES:

- 110 V A/C Adapter Regulator Mag Card Rdr#
Printer Paper Printer Ribbon Casio #
Mouthpieces Carrying Case Dry Gas

Other (Please Specify) Warranty Exp. Date Oct 2013

Repair Information:

Test #

Table with 4 columns: Part Number, Description, Qty, Total Cost. Rows include Motor, CAL, and Labor.

Repair Notes:

FC# ARBL-1170
Replaced motor, because it made a squeaking noise.
CAL w/ DC & Ops check

Service Technician: [Signature]

Date: 2/10/2011