

# Return & Repair Form:

## Customer Information:

Company Name: Mansfield Corp B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3-4-11 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WA - 0193  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>12098</u>	<u>Washer</u>	<u>2</u>	<u>N/C W</u>
<u>12094</u>	<u>screw</u>	<u>2</u>	<u>N/C W</u>
<u>12095</u>	<u>NUT</u>	<u>2</u>	<u>N/C W</u>
<u>12008</u>	<u>Bumper</u>	<u>2</u>	<u>N/C W</u>
<u>M-3c part</u>	<u>7110 Case</u>	<u>1</u>	<u>N/C W</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

### Repair Notes:

Replace Damage 7110 Case.  
Cal w/DC ; ops check

Service Technician: [Signature]

Date: 3/7/2011