

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: New Jersey State Police - Buena Vista Station, Buena Vista Twp., NJ

Date Received: 06/27/14 Date given to service: 06/27/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810 ☒ 7110 ☐ 8610 ☐ 7510 ☐ DT5000 ☐ 7410 Upper-half ☐ 9510 ☐ 7410 Whole

Serial Number: AR WC-0007
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: Jan. 2015

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	4	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
Misc. Part	LCD Display (Cabbage Part)	1	NC/W

Repair Notes: Replaced defective fuel cell, spacer plates, C156 and C167.
C156 and C167 were changed due to issues connecting to the modem.

Fuel Cell = ARFD-2199

CAL W/QC AND OPS CHECK

Service Technician BC Date: 07/08/14