

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: NJSP-ADTU, Hamilton, NJ

Date Received: 02/15/2016

Date given to service: 02/15/2016

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☒ 3DAY

☐ 2DAY

☐ NDA-PRI

☐ NDA-STD

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR WD-0107

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defective fuel cell.

Ran multiple tests to reproduce the unit getting stuck in searching mode when trying to print a copy but was unable to duplicate.

Fuel Cell = ARHM-0718

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/01/2016