

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: NJSP - ADTU, Hamilton, NJ

Date Received: 03/18/2016

Date given to service: 03/18/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WD-0107

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC
4414166	Labor	5	NC
8315075	Motherboard	1	NC

Repair Notes: Replaced defective motherboard.

Second time unit was in for the same issue, changing motherboard will resolve this issue.

CAL W/QC AND OPS CHECK

Service Technician _____

Date: 03/18/2016