

Dräger**RETURN AND REPAIR FORM****Customer Information**

B:

S:

Company Name: City of Lambertville P.D., Lambertville, NJDate Received: 01/10/13 Date given to service: 01/10/13Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STDProduct: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 WholeSerial Number: AR WE-0003Printer Serial#: ARSim Serial#: DDProbe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In**Accessories**☐ 110V A/C Adapter
☐ Mouthpieces
☐ Dry Gas☐ Regulator
☐ 9510 Stylus
☐ Other (please specify) _____☐ Printer Ribbon
☐ 9510 Top Cover☐ Printer Paper
☐ Carrying Case**Repair Information:****Test#:**

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel cell	1	NC/W
6808486	Spacer Plate	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Replaced defective fuel cell and spacer plate.Fuel Cell ARZL-1985 by TMReplaced Ceramic Capacitor = CrackedReplaced Cermetek = Modern would not answerCAL W/QC AND OPS CHECKService Technician: BC/TMDate: 1/29/13