

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Washington Twp. Police, Washington, NJ

Date Received: 02/21/14

Date given to service: 02/21/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole
Serial Number: AR WE-0004
Printer Serial#: AR
Sim Serial#: DD
Probe Serial#: DD

Warranty Expires: October 2015

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	2	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6809458	IR Detector	1	NC/W

Repair Notes: Replaced IR Detector, fuel cell, and spacer plates.

Ir Detector was replaced due to unit giving BLOWING NOT ALLOWED error.

Fuel Cell = AREL-1310

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/26/14