

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Bordentown City P.D., Bordentown, NJ

Date Received: 06/19/14

Date given to service: 06/19/14

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☐ 3DAY

☐ 2DAY

☒ NDA-PRI

☐ NDA-STD

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR WE-0007

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	9	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
12003	Hose Barb 1/8" 7110 Outlet	1	NC/W
8315075	Alcotest 7110 Motherboard MKII	1	NC/W

Repair Notes: Replaced defective motherboard, hosebarb, fuel cell, and spacer plates.

Motherboard was giving wrong wet bath readings.

Fuel Cell = ARFD-1854

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 06/30/14