

Dräger safety

Return & Repair Form

Customer Information:

Customer: Lopateang Police Dept. Cust # 58568
 Date of Return: 9/5/07 Carrier: Fedex Method: Priority Overnight

Product Information:

Model: 7110 7410 6510 6810Serial # ARWE-0009

Description: A B PLUS, SCREENER

Printer Serial #

Whole Inst. Top Half

Simulator Serial #

ER:

Probe Serial #

Unit ☐ Returned to stock on:

Accessories: (check all that apply):

☐ 110VAC Adapter☐ Regulator☐ Mag Card Rdr #☐ Printer Paper☐ Printer Ribbon☐ Casio #☐ Mouthpieces☐ Carrying Case☐ Dry Gas☐ Other (specify):

Warranty Expiration Date:

Test Information:

Test #

Reason for Return: LOW EC RESULTS

Part #	Description	Qty	Total
6808455	FUEL CELL	1	WARRANTY
MPCAC 71	7110 CAP	1	WARR.
MPLABOR	LABOR		WARR.

Repair Notes: REPLACE FC / CAP & OPS CHECKS

Technician:

Date: 09-06-07