

Dräger safety

Internal Use Only

Return & Repair Form

Customer Information:

CUSTOMER: SPRINGFIELD TRIP Cust #
Date of Return: 8-6-07 Carrier: FOO Method: STD OVERNITE

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWIE-CD13
Description: A B PLUS SCREENER _ Printer Serial #
Whole Inst. Top Half Simulator Serial #
OTHER: Probe Serial #

Demo Unit ☐ Returned to stock on:

Accessories: (check all that apply):

- ☐ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr #
☐ Printer Paper ☐ Printer Ribbon ☐ Casio #
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
☐ Other (specify):

Warranty Expiration Date:

Repair Information:

Test #

Reason for Return: EC RESULT PROBLEM

Part #	Description	Qty	Total
6808455	FUEL CELL	1	WARRANTY
MPCAL 71	CAL 7110	1	WARRANTY
MPLABOR	LABOR	.5	WARRANTY

Repair Notes: REPLACED FC - SLIGHT CORROSION BY LEADS / NEW = ARJD-1603 /
CAL - QC & OPS CHECKS

Technician:

Date: 08-06-07