

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Springfield Twp P.D., Johnston, VT

Date Received: 10/11/11 Date given to service: _____

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WE-0012

Printer Serial#: AR _____

Sim Serial#: DD _____

Probe Serial#: DD _____

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
8315075	motor board	1	n/c w
mp cal 71	calibration	1	n/c w
mp Labor	Labor	.5	n/c w

Repair Notes: Unit would not come out of stand-by mode, installed new motor board.

Cal w/DC, ops check

Service Technician JS Date: 12-21-11