

Return & Repair Form

Customer Information:

CUSTOMER Beach Haven Police Dep. Cust # _____

Date of Return: 6-4-7 Carrier: FEDEX Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810

Serial # ARWE-0013

Description: A B PLUS SCREENER

Printer Serial # _____

Whole Inst. Top Half

Simulator Serial # _____

OTHER: _____

Probe Serial # _____

Demo Unit ☐ Returned to stock on: _____

** NOT PACKAGED GOOD.*

Accessories: (check all that apply):

☐ 110VAC Adapter

☐ Regulator

☐ Mag Card Rdr # _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

☐ Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC Problem

Part #	Description	Qty	Total
<u>0808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MP CAL 71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced Questionable FLC. Full
cal ops check + Q.C.

Technician: CRD

Date: 6/5/07