

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Beach Haven P.D., Beach Haven N.J.

Date Received: 09/12/11 Date given to service: _____

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WE-0013
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
680 8455	Fuel Cell	1	N/C W
680 8486	Plate	1	N/C W
mp ccl 71	Calibrate	1	N/C W
mp Labor	Labor	.5	N/C W

Repair Notes: Changed Fuel cell AR 19-0369, because of
Low EC reading.

Cal w/DC / ops check

Service Technician: [Signature]

Date: 9-14-11 10466712

58982-3-29-11