

Return & Repair Form:

Customer Information:

Company Name: New Jersey State Police B: _____ S: _____

Date Received: 7-14-09 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR WE - 0014

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: _____

Other _____

Probe Ser # DD P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

PCB, FC

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C w/arr</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w/arr</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C w/arr</u>
<u>MP cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w/arr</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C w/arr</u>

Repair Notes:

Replace PCB, FC

Cal w/ QC / Ops check

Service Technician: JS

Date: 07/15/2009