

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Dehanco, P.D. DeLanco NJ

Date Received: 03/09/12 Date given to service: 03/09/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☒ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole
 Serial Number: AR WE-0015
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/C W
mp cal 71	Calibration	1	N/C W
mp Labor	Labor	.5	N/C W

Repair Notes: Installed new capacitor at Location C156. also
Installed a Hose from the pump to the solenoid, because
of a Pinch in the Hose.
Install new Fuel Cell ARCK-1272 to replace busted
Fuel Cell ARAJ-0169. Unit also would not complete
a solution change.
Cal w/ QC, Ops check

Service Technician: ps Date: 3-14-12