

Return & Repair Form:

Customer Information:

Company Name: NJ State Police B: _____ S: _____

Date Received: _____ Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: ARWE - 0018

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: _____

Other _____

Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

FL

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE FC + 1 PLATE

CAL W/QC & Ops Check

Service Technician: JS

Date: 7-8-09