

Return & Repair Form:

Customer Information:

Company Name: Seaside Park B: _____ S: _____
 Date Received: 4-12-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR WE 0020
 9510-Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/CW</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/CW</u>
<u>Mix Part</u>	<u>PRESSURE SENSOR</u>	<u>1</u>	<u>N/CW</u>
<u>MP CALOR</u>	<u>CALOR</u>	<u>.5</u>	<u>N/CW</u>

Repair Notes:

NEW FC# ARAN-1049

Cal w/DC & OPS Check

Re-aligned Part Pressure sensor, unit would NOT do a ABA test through the simu layer.

Service Technician:

[Signature]

Date:

04-12-2010