

Dräger safety

Return & Repair Form

Internal Use Only

RETURN
P1

Customer Information:

CUSTOMER: Hillside PD

Cust #

Date of Return: 8/2/07

Carrier: Fedex

Method: P1

Product Information:

Product: 7110 7410 6510 6810

Serial # ARWE 0023

Description: A B PLUS SCREENER

Printer Serial #

Whole Inst. Top Half

Simulator Serial #

OTHER:

Probe Serial #

Demo Unit ☐ Returned to stock on:

Accessories: (check all that apply):

☐ 110VAC Adapter

☐ Regulator

☐ Mag Card Rdr #

☐ Printer Paper

☐ Printer Ribbon

☐ Casio #

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

☒ Other (specify): Power Cord

Warranty Expiration Date:

Repair Information:

Test #

Reason for Return: ERROR 32

Part #	Description	Qty	Total
6810269	MOTOR	1	N/C
6807819	SWITCH	1	N/C
MPCAL 71	7110 CAL	1	N/C
MPLABOR		.5	N/C

Repair Notes: DID NOT DUPLICATE ERROR- REPLACED MOTOR/
SWITCH / CAL & OPS CHECKS

Technician: AE

Date: 08-02-07