

RETURN AND REPAIR FORM

Dräger

Customer Information

B: #

S: #

Company Name: NJSP - Hamilton Technology Center, Hamilton, NJ

Date Received: 02/14/2017

Date given to service: 02/14/2017

Carrier: ☐ FedEx ☒ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☒ NDA-STD

Product: ☐ 6510 ☐ 6810 ☒ 7110 ☐ 8610 ☐ 7510 ☐ DT5000 ☐ 7410 Upper-half ☐ 9510 ☐ 7410 Whole

Serial Number: AR WE-0037
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARJL-0027

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/09/17