

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: New Jersey State Police, Roxbury, NJ

Date Received: 05/10/2016

Date given to service: 05/10/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WE-0038

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161 | Calibration | 1 | |
| 4414166 | Labor | .5 | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes: Performed and passed all QC.

Irv was too low out of range which can cause Error 23, this was adjusted to be in range.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 06/03/2016