

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: NJSP - Netcong, Roxbury, NJ

Date Received: 12/07/2016

Date given to service: 12/07/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WE-0039

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	6	NC/W
12130	Modem Module	1	NC/W
6809514	Breath Hose	1	NC/W
8315075	7110 Motherboard	1	NC/W

Repair Notes: Replaced defective modem module, breath hose, and motherboard.

Breath hose was changed due to Error 72, Motherboard was changed due to the unit freezing up when orange button is pressed, and modem module was changed due to modem issues.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 12/12/2016