

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: New Jersey State Police - Tuckerton Station, Eagleswood, NJ

Date Received: 02/12/2016

Date given to service: 02/12/2016

Carrier: ☐ FedEx ☒ UPS ☐ USPS Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR WE-0080

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6808374	Geared Engine (Motor)	1	NC/W

Repair Notes: Replaced defective geared engine, fuel cell, and spacer plates.

Geared engine was changed due to grinding.

Fuel Cell = ARHM-0821

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/25/2016