

# Return & Repair Form:

## Customer Information:

Company Name: Lawrence PD B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 4/18/2011 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-~~7110~~-6510-6810-7510 Serial #: AR WF - 0359  
 9510-Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>w/c w</u>
<u>6808480</u>	<u>Plate</u>	<u>1</u>	<u>w/c w</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>w/c w</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c w</u>

## Repair Notes:

FC# ARCA-1090  
Cal w/PC & ops check

Service Technician: AS

Date: 4/19/2011