

Return & Repair Form

Customer Information:

CUSTOMER Lacey Trust Cust # 59370
59371
 Date of Return: 12/5/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: _____

Product Information:

Product: 7110 7410 6510 6810 Serial # AR WF - 0366
 Description: A B PLUS SCREENER _ Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # DD P -
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Reason for Return: ~~Catalytic~~ FUEL CELL PROBLEM Test # _____

Part #	Description	Qty	Total
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPCAL71</u>	<u>CAC</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR N/C</u>

Repair Notes: REPLACE FC / CAC w/ QC & OPS CHECKS

Technician: [Signature]

Date: 12-10-07