

## RETURN AND REPAIR FORM

### Customer Information

B:

S:

Company Name: Ship Bottom PD.

Date Received: 7-28-11 Date given to service: 7-28-11

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WF-0396

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

### Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

Test#: \_\_\_\_\_

### Repair Information

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
6809514	Breath hose	1	n/cw
8100902	Flow Sensor	1	n/cw
mp cal 71	Calibrate	1	n/cw
mp labor	Labor	.5	Labor

Repair Notes: Cleaned spots off a cuvette, also replaced broken flow sensor, breath hose was inspected and one of the solder joints looked rough, so I replaced the breath hose for preventive maintenance.

cal w/ DC & ops check

Service Technician: [Signature] Date: 8-2-11