

# Return & Repair Form:

## Customer Information:

Company Name: Peterson PD B: 65011 S: \_\_\_\_\_  
 Date Received: 3 Jan 11 Date Given to Service: 3 Jan 11  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: ARWF - 0397  
9510-Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date May 2013

## Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C w</u>
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w</u>
<u>mp labor</u>	<u>labor</u>	<u>1.5</u>	<u>N/C w</u>

Repair Notes:

FC# ARCA - 0771

cal w/DC; ops check

Service Technician: JS

Date: 6-7-11