

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Highlands Boro Police Dept. , Highlands ,NJ

Date Received: 07/06/12

Date given to service: 07/06/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WF-0400

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
Misc part	(PCB) 8315075	1	NC/W
MP Labor	Labor	1	NC/W

Repair Notes: Installed PCB ,because unit would not reset with orange button.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 07-12-12