

Return & Repair Form:

Customer Information:

Company Name: P. de Brach B: _____ S: _____

Date Received: 4-24-09 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR 403 - 0003

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top ½ Sim Ser #: _____

Other _____ Probe Ser # DD P -

ACCESSORIES:

☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____

☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____

☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

Replace FC
CAL w/ QC & ops check

Service Technician: JS

Date: 04/24/2009